

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190102-ICV1	Arsenic KED ICPMS 200.8	40.41	ug/L	101.0%	1/7/2019 14:41	40.00	ug/L
S190102-ICV1	Cadmium KED ICPMS 200.8	80.56	ug/L	101.0%	1/7/2019 14:41	80.00	ug/L
S190102-ICV1	Chromium KED ICPMS 200.8	239.39	ug/L	99.7%	1/7/2019 14:41	240.00	ug/L
S190102-ICV1	Copper KED ICPMS 200.8	82.11	ug/L	103.0%	1/7/2019 14:41	80.00	ug/L
S190102-ICV1	Lead KED ICPMS 200.8	41.39	ug/L	103.0%	1/7/2019 14:41	40.00	ug/L
S190102-ICV1	Nickel KED ICPMS 200.8	81.25	ug/L	102.0%	1/7/2019 14:41	80.00	ug/L
S190102-ICV1	Silver KED ICPMS 200.8	40.70	ug/L	102.0%	1/7/2019 14:41	40.00	ug/L
S190102-ICV1	Zinc KED ICPMS 200.8	165.04	ug/L	103.0%	1/7/2019 14:41	160.00	ug/L
S190102-CCV1	Arsenic KED ICPMS 200.8	51.14	ug/L	102.0%	1/7/2019 15:05	50.00	ug/L
S190102-CCV1	Cadmium KED ICPMS 200.8	100.95	ug/L	101.0%	1/7/2019 15:05	100.00	ug/L
S190102-CCV1	Chromium KED ICPMS 200.8	305.35	ug/L	102.0%	1/7/2019 15:05	300.00	ug/L
S190102-CCV1	Copper KED ICPMS 200.8	103.02	ug/L	103.0%	1/7/2019 15:05	100.00	ug/L
S190102-CCV1	Lead KED ICPMS 200.8	50.44	ug/L	101.0%	1/7/2019 15:05	50.00	ug/L
S190102-CCV1	Nickel KED ICPMS 200.8	102.41	ug/L	102.0%	1/7/2019 15:05	100.00	ug/L
S190102-CCV1	Silver KED ICPMS 200.8	50.51	ug/L	101.0%	1/7/2019 15:05	50.00	ug/L
S190102-CCV1	Zinc KED ICPMS 200.8	207.86	ug/L	104.0%	1/7/2019 15:05	200.00	ug/L
S190102-CCV2	Arsenic KED ICPMS 200.8	51.37	ug/L	103.0%	1/7/2019 16:14	50.00	ug/L
S190102-CCV2	Cadmium KED ICPMS 200.8	100.01	ug/L	100.0%	1/7/2019 16:14	100.00	ug/L
S190102-CCV2	Chromium KED ICPMS 200.8	304.07	ug/L	101.0%	1/7/2019 16:14	300.00	ug/L
S190102-CCV2	Copper KED ICPMS 200.8	102.52	ug/L	103.0%	1/7/2019 16:14	100.00	ug/L
S190102-CCV2	Lead KED ICPMS 200.8	49.80	ug/L	99.6%	1/7/2019 16:14	50.00	ug/L
S190102-CCV2	Nickel KED ICPMS 200.8	102.12	ug/L	102.0%	1/7/2019 16:14	100.00	ug/L
S190102-CCV2	Silver KED ICPMS 200.8	50.23	ug/L	100.0%	1/7/2019 16:14	50.00	ug/L
S190102-CCV2	Zinc KED ICPMS 200.8	207.32	ug/L	104.0%	1/7/2019 16:14	200.00	ug/L
S190102-CCV3	Arsenic KED ICPMS 200.8	51.56	ug/L	103.0%	1/7/2019 17:27	50.00	ug/L
S190102-CCV3	Cadmium KED ICPMS 200.8	101.61	ug/L	102.0%	1/7/2019 17:27	100.00	ug/L
S190102-CCV3	Chromium KED ICPMS 200.8	310.86	ug/L	104.0%	1/7/2019 17:27	300.00	ug/L
S190102-CCV3	Copper KED ICPMS 200.8	103.94	ug/L	104.0%	1/7/2019 17:27	100.00	ug/L
S190102-CCV3	Lead KED ICPMS 200.8	49.77	ug/L	99.5%	1/7/2019 17:27	50.00	ug/L
S190102-CCV3	Nickel KED ICPMS 200.8	103.34	ug/L	103.0%	1/7/2019 17:27	100.00	ug/L
S190102-CCV3	Silver KED ICPMS 200.8	50.66	ug/L	101.0%	1/7/2019 17:27	50.00	ug/L
S190102-CCV3	Zinc KED ICPMS 200.8	212.14	ug/L	106.0%	1/7/2019 17:27	200.00	ug/L
S190102-CCV4	Arsenic KED ICPMS 200.8	50.95	ug/L	102.0%	1/7/2019 18:35	50.00	ug/L
S190102-CCV4	Cadmium KED ICPMS 200.8	101.72	ug/L	102.0%	1/7/2019 18:35	100.00	ug/L
S190102-CCV4	Chromium KED ICPMS 200.8	306.53	ug/L	102.0%	1/7/2019 18:35	300.00	ug/L
S190102-CCV4	Copper KED ICPMS 200.8	101.88	ug/L	102.0%	1/7/2019 18:35	100.00	ug/L
S190102-CCV4	Lead KED ICPMS 200.8	49.12	ug/L	98.2%	1/7/2019 18:35	50.00	ug/L
S190102-CCV4	Nickel KED ICPMS 200.8	101.75	ug/L	102.0%	1/7/2019 18:35	100.00	ug/L
S190102-CCV4	Silver KED ICPMS 200.8	50.39	ug/L	101.0%	1/7/2019 18:35	50.00	ug/L
S190102-CCV4	Zinc KED ICPMS 200.8	208.85	ug/L	104.0%	1/7/2019 18:35	200.00	ug/L
S190102-CCV5	Arsenic KED ICPMS 200.8	51.40	ug/L	103.0%	1/7/2019 19:44	50.00	ug/L
S190102-CCV5	Cadmium KED ICPMS 200.8	102.02	ug/L	102.0%	1/7/2019 19:44	100.00	ug/L
S190102-CCV5	Chromium KED ICPMS 200.8	305.95	ug/L	102.0%	1/7/2019 19:44	300.00	ug/L
S190102-CCV5	Copper KED ICPMS 200.8	102.38	ug/L	102.0%	1/7/2019 19:44	100.00	ug/L
S190102-CCV5	Lead KED ICPMS 200.8	48.65	ug/L	97.3%	1/7/2019 19:44	50.00	ug/L
S190102-CCV5	Nickel KED ICPMS 200.8	101.68	ug/L	102.0%	1/7/2019 19:44	100.00	ug/L
S190102-CCV5	Silver KED ICPMS 200.8	50.39	ug/L	101.0%	1/7/2019 19:44	50.00	ug/L
S190102-CCV5	Zinc KED ICPMS 200.8	210.54	ug/L	105.0%	1/7/2019 19:44	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190103-ICV1	Arsenic KED ICPMS 200.8	40.86	ug/L	102.00%	1/8/2019 14:06	40.00	ug/L
S190103-ICV1	Cadmium KED ICPMS 200.8	81.53	ug/L	102.00%	1/8/2019 14:06	80.00	ug/L
S190103-ICV1	Chromium KED ICPMS 200.8	243.23	ug/L	101.00%	1/8/2019 14:06	240.00	ug/L
S190103-ICV1	Copper KED ICPMS 200.8	82.59	ug/L	103.00%	1/8/2019 14:06	80.00	ug/L
S190103-ICV1	Lead KED ICPMS 200.8	41.81	ug/L	105.00%	1/8/2019 14:06	40.00	ug/L
S190103-ICV1	Nickel KED ICPMS 200.8	82.11	ug/L	103.00%	1/8/2019 14:06	80.00	ug/L
S190103-ICV1	Silver KED ICPMS 200.8	40.72	ug/L	102.00%	1/8/2019 14:06	40.00	ug/L
S190103-ICV1	Zinc KED ICPMS 200.8	167.21	ug/L	105.00%	1/8/2019 14:06	160.00	ug/L
S190103-CCV1	Arsenic KED ICPMS 200.8	50.89	ug/L	102.00%	1/8/2019 14:30	50.00	ug/L
S190103-CCV1	Cadmium KED ICPMS 200.8	99.88	ug/L	99.90%	1/8/2019 14:30	100.00	ug/L
S190103-CCV1	Chromium KED ICPMS 200.8	303.77	ug/L	101.00%	1/8/2019 14:30	300.00	ug/L
S190103-CCV1	Copper KED ICPMS 200.8	101.45	ug/L	101.00%	1/8/2019 14:30	100.00	ug/L
S190103-CCV1	Lead KED ICPMS 200.8	50.63	ug/L	101.00%	1/8/2019 14:30	50.00	ug/L
S190103-CCV1	Nickel KED ICPMS 200.8	101.39	ug/L	101.00%	1/8/2019 14:30	100.00	ug/L
S190103-CCV1	Silver KED ICPMS 200.8	49.79	ug/L	99.60%	1/8/2019 14:30	50.00	ug/L
S190103-CCV1	Zinc KED ICPMS 200.8	206.84	ug/L	103.00%	1/8/2019 14:30	200.00	ug/L
S190103-CCV2	Arsenic KED ICPMS 200.8	50.91	ug/L	102.00%	1/8/2019 16:01	50.00	ug/L
S190103-CCV2	Cadmium KED ICPMS 200.8	99.47	ug/L	99.50%	1/8/2019 16:01	100.00	ug/L
S190103-CCV2	Chromium KED ICPMS 200.8	307.48	ug/L	102.00%	1/8/2019 16:01	300.00	ug/L
S190103-CCV2	Copper KED ICPMS 200.8	101.04	ug/L	101.00%	1/8/2019 16:01	100.00	ug/L
S190103-CCV2	Lead KED ICPMS 200.8	49.96	ug/L	99.90%	1/8/2019 16:01	50.00	ug/L
S190103-CCV2	Nickel KED ICPMS 200.8	101.10	ug/L	101.00%	1/8/2019 16:01	100.00	ug/L
S190103-CCV2	Silver KED ICPMS 200.8	48.38	ug/L	96.80%	1/8/2019 16:01	50.00	ug/L
S190103-CCV2	Zinc KED ICPMS 200.8	210.08	ug/L	105.00%	1/8/2019 16:01	200.00	ug/L
S190103-CCV3	Arsenic KED ICPMS 200.8	50.19	ug/L	100.00%	1/8/2019 17:13	50.00	ug/L
S190103-CCV3	Cadmium KED ICPMS 200.8	99.17	ug/L	99.20%	1/8/2019 17:13	100.00	ug/L
S190103-CCV3	Chromium KED ICPMS 200.8	306.05	ug/L	102.00%	1/8/2019 17:13	300.00	ug/L
S190103-CCV3	Copper KED ICPMS 200.8	99.47	ug/L	99.50%	1/8/2019 17:13	100.00	ug/L
S190103-CCV3	Lead KED ICPMS 200.8	50.15	ug/L	100.00%	1/8/2019 17:13	50.00	ug/L
S190103-CCV3	Nickel KED ICPMS 200.8	99.75	ug/L	99.70%	1/8/2019 17:13	100.00	ug/L
S190103-CCV3	Silver KED ICPMS 200.8	48.23	ug/L	96.50%	1/8/2019 17:13	50.00	ug/L
S190103-CCV3	Zinc KED ICPMS 200.8	206.75	ug/L	103.00%	1/8/2019 17:13	200.00	ug/L
S190103-CCV4	Arsenic KED ICPMS 200.8	50.35	ug/L	101.00%	1/8/2019 18:30	50.00	ug/L
S190103-CCV4	Cadmium KED ICPMS 200.8	98.46	ug/L	98.50%	1/8/2019 18:30	100.00	ug/L
S190103-CCV4	Chromium KED ICPMS 200.8	307.62	ug/L	103.00%	1/8/2019 18:30	300.00	ug/L
S190103-CCV4	Copper KED ICPMS 200.8	99.10	ug/L	99.10%	1/8/2019 18:30	100.00	ug/L
S190103-CCV4	Lead KED ICPMS 200.8	50.56	ug/L	101.00%	1/8/2019 18:30	50.00	ug/L
S190103-CCV4	Nickel KED ICPMS 200.8	99.79	ug/L	99.80%	1/8/2019 18:30	100.00	ug/L
S190103-CCV4	Silver KED ICPMS 200.8	47.56	ug/L	95.10%	1/8/2019 18:30	50.00	ug/L
S190103-CCV4	Zinc KED ICPMS 200.8	208.36	ug/L	104.00%	1/8/2019 18:30	200.00	ug/L
S190103-CCV5	Arsenic KED ICPMS 200.8	49.92	ug/L	99.80%	1/8/2019 19:42	50.00	ug/L
S190103-CCV5	Cadmium KED ICPMS 200.8	99.25	ug/L	99.30%	1/8/2019 19:42	100.00	ug/L
S190103-CCV5	Chromium KED ICPMS 200.8	308.53	ug/L	103.00%	1/8/2019 19:42	300.00	ug/L
S190103-CCV5	Copper KED ICPMS 200.8	99.46	ug/L	99.50%	1/8/2019 19:42	100.00	ug/L
S190103-CCV5	Lead KED ICPMS 200.8	51.08	ug/L	102.00%	1/8/2019 19:42	50.00	ug/L
S190103-CCV5	Nickel KED ICPMS 200.8	99.68	ug/L	99.70%	1/8/2019 19:42	100.00	ug/L
S190103-CCV5	Silver KED ICPMS 200.8	47.32	ug/L	94.60%	1/8/2019 19:42	50.00	ug/L
S190103-CCV5	Zinc KED ICPMS 200.8	209.32	ug/L	105.00%	1/8/2019 19:42	200.00	ug/L

S190103-CCV6	Arsenic KED ICPMS 200.8	50.27	ug/L	101.00%	1/8/2019 20:49	50.00	ug/L
S190103-CCV6	Cadmium KED ICPMS 200.8	99.31	ug/L	99.30%	1/8/2019 20:49	100.00	ug/L
S190103-CCV6	Chromium KED ICPMS 200.8	310.85	ug/L	104.00%	1/8/2019 20:49	300.00	ug/L
S190103-CCV6	Copper KED ICPMS 200.8	100.15	ug/L	100.00%	1/8/2019 20:49	100.00	ug/L
S190103-CCV6	Lead KED ICPMS 200.8	51.24	ug/L	102.00%	1/8/2019 20:49	50.00	ug/L
S190103-CCV6	Nickel KED ICPMS 200.8	100.36	ug/L	100.00%	1/8/2019 20:49	100.00	ug/L
S190103-CCV6	Silver KED ICPMS 200.8	47.13	ug/L	94.30%	1/8/2019 20:49	50.00	ug/L
S190103-CCV6	Zinc KED ICPMS 200.8	210.48	ug/L	105.00%	1/8/2019 20:49	200.00	ug/L
S190103-CCV7	Arsenic KED ICPMS 200.8	50.28	ug/L	101.00%	1/8/2019 22:07	50.00	ug/L
S190103-CCV7	Cadmium KED ICPMS 200.8	98.31	ug/L	98.30%	1/8/2019 22:07	100.00	ug/L
S190103-CCV7	Chromium KED ICPMS 200.8	309.19	ug/L	103.00%	1/8/2019 22:07	300.00	ug/L
S190103-CCV7	Copper KED ICPMS 200.8	98.82	ug/L	98.80%	1/8/2019 22:07	100.00	ug/L
S190103-CCV7	Lead KED ICPMS 200.8	51.43	ug/L	103.00%	1/8/2019 22:07	50.00	ug/L
S190103-CCV7	Nickel KED ICPMS 200.8	99.63	ug/L	99.60%	1/8/2019 22:07	100.00	ug/L
S190103-CCV7	Silver KED ICPMS 200.8	46.44	ug/L	92.90%	1/8/2019 22:07	50.00	ug/L
S190103-CCV7	Zinc KED ICPMS 200.8	210.58	ug/L	105.00%	1/8/2019 22:07	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190105-ICV1	Arsenic KED ICPMS 200.8	41.31	ug/L	103.00%	1/14/2019 17:31	40.00	ug/L
S190105-ICV1	Cadmium KED ICPMS 200.8	80.13	ug/L	100.00%	1/14/2019 17:31	80.00	ug/L
S190105-ICV1	Chromium KED ICPMS 200.8	242.14	ug/L	101.00%	1/14/2019 17:31	240.00	ug/L
S190105-ICV1	Copper KED ICPMS 200.8	83.35	ug/L	104.00%	1/14/2019 17:31	80.00	ug/L
S190105-ICV1	Lead KED ICPMS 200.8	41.61	ug/L	104.00%	1/14/2019 17:31	40.00	ug/L
S190105-ICV1	Nickel KED ICPMS 200.8	82.32	ug/L	103.00%	1/14/2019 17:31	80.00	ug/L
S190105-ICV1	Silver KED ICPMS 200.8	39.74	ug/L	99.40%	1/14/2019 17:31	40.00	ug/L
S190105-ICV1	Zinc KED ICPMS 200.8	166.97	ug/L	104.00%	1/14/2019 17:31	160.00	ug/L
S190105-CCV1	Arsenic KED ICPMS 200.8	50.94	ug/L	102.00%	1/14/2019 17:55	50.00	ug/L
S190105-CCV1	Cadmium KED ICPMS 200.8	100.89	ug/L	101.00%	1/14/2019 17:55	100.00	ug/L
S190105-CCV1	Chromium KED ICPMS 200.8	305.10	ug/L	102.00%	1/14/2019 17:55	300.00	ug/L
S190105-CCV1	Copper KED ICPMS 200.8	103.24	ug/L	103.00%	1/14/2019 17:55	100.00	ug/L
S190105-CCV1	Lead KED ICPMS 200.8	50.26	ug/L	101.00%	1/14/2019 17:55	50.00	ug/L
S190105-CCV1	Nickel KED ICPMS 200.8	102.25	ug/L	102.00%	1/14/2019 17:55	100.00	ug/L
S190105-CCV1	Silver KED ICPMS 200.8	50.04	ug/L	100.00%	1/14/2019 17:55	50.00	ug/L
S190105-CCV1	Zinc KED ICPMS 200.8	208.79	ug/L	104.00%	1/14/2019 17:55	200.00	ug/L
S190105-CCV2	Arsenic KED ICPMS 200.8	51.25	ug/L	102.00%	1/14/2019 18:59	50.00	ug/L
S190105-CCV2	Cadmium KED ICPMS 200.8	100.97	ug/L	101.00%	1/14/2019 18:59	100.00	ug/L
S190105-CCV2	Chromium KED ICPMS 200.8	306.87	ug/L	102.00%	1/14/2019 18:59	300.00	ug/L
S190105-CCV2	Copper KED ICPMS 200.8	103.20	ug/L	103.00%	1/14/2019 18:59	100.00	ug/L
S190105-CCV2	Lead KED ICPMS 200.8	50.01	ug/L	100.00%	1/14/2019 18:59	50.00	ug/L
S190105-CCV2	Nickel KED ICPMS 200.8	102.51	ug/L	103.00%	1/14/2019 18:59	100.00	ug/L
S190105-CCV2	Silver KED ICPMS 200.8	50.29	ug/L	101.00%	1/14/2019 18:59	50.00	ug/L
S190105-CCV2	Zinc KED ICPMS 200.8	208.27	ug/L	104.00%	1/14/2019 18:59	200.00	ug/L
S190105-CCV3	Arsenic KED ICPMS 200.8	51.20	ug/L	102.00%	1/14/2019 20:09	50.00	ug/L
S190105-CCV3	Cadmium KED ICPMS 200.8	101.15	ug/L	101.00%	1/14/2019 20:09	100.00	ug/L
S190105-CCV3	Chromium KED ICPMS 200.8	301.92	ug/L	101.00%	1/14/2019 20:09	300.00	ug/L
S190105-CCV3	Copper KED ICPMS 200.8	102.40	ug/L	102.00%	1/14/2019 20:09	100.00	ug/L
S190105-CCV3	Lead KED ICPMS 200.8	49.75	ug/L	99.50%	1/14/2019 20:09	50.00	ug/L
S190105-CCV3	Nickel KED ICPMS 200.8	102.36	ug/L	102.00%	1/14/2019 20:09	100.00	ug/L
S190105-CCV3	Silver KED ICPMS 200.8	49.94	ug/L	99.90%	1/14/2019 20:09	50.00	ug/L
S190105-CCV3	Zinc KED ICPMS 200.8	206.08	ug/L	103.00%	1/14/2019 20:09	200.00	ug/L
S190105-CCV4	Arsenic KED ICPMS 200.8	51.36	ug/L	103.00%	1/14/2019 21:17	50.00	ug/L
S190105-CCV4	Cadmium KED ICPMS 200.8	101.43	ug/L	101.00%	1/14/2019 21:17	100.00	ug/L
S190105-CCV4	Chromium KED ICPMS 200.8	306.06	ug/L	102.00%	1/14/2019 21:17	300.00	ug/L
S190105-CCV4	Copper KED ICPMS 200.8	103.83	ug/L	104.00%	1/14/2019 21:17	100.00	ug/L
S190105-CCV4	Lead KED ICPMS 200.8	50.97	ug/L	102.00%	1/14/2019 21:17	50.00	ug/L
S190105-CCV4	Nickel KED ICPMS 200.8	103.67	ug/L	104.00%	1/14/2019 21:17	100.00	ug/L
S190105-CCV4	Silver KED ICPMS 200.8	50.32	ug/L	101.00%	1/14/2019 21:17	50.00	ug/L
S190105-CCV4	Zinc KED ICPMS 200.8	209.23	ug/L	105.00%	1/14/2019 21:17	200.00	ug/L
S190105-CCV5	Arsenic KED ICPMS 200.8	51.73	ug/L	103.00%	1/14/2019 22:24	50.00	ug/L
S190105-CCV5	Cadmium KED ICPMS 200.8	101.05	ug/L	101.00%	1/14/2019 22:24	100.00	ug/L
S190105-CCV5	Chromium KED ICPMS 200.8	302.72	ug/L	101.00%	1/14/2019 22:24	300.00	ug/L
S190105-CCV5	Copper KED ICPMS 200.8	102.75	ug/L	103.00%	1/14/2019 22:24	100.00	ug/L
S190105-CCV5	Lead KED ICPMS 200.8	50.41	ug/L	101.00%	1/14/2019 22:24	50.00	ug/L
S190105-CCV5	Nickel KED ICPMS 200.8	103.36	ug/L	103.00%	1/14/2019 22:24	100.00	ug/L
S190105-CCV5	Silver KED ICPMS 200.8	50.10	ug/L	100.00%	1/14/2019 22:24	50.00	ug/L

S190105-CCV5	Zinc KED ICPMS 200.8	207.81	ug/L	104.00%	1/14/2019 22:24	200.00	ug/L
S190105-CCV6	Arsenic KED ICPMS 200.8	51.54	ug/L	103.00%	1/14/2019 22:51	50.00	ug/L
S190105-CCV6	Cadmium KED ICPMS 200.8	102.08	ug/L	102.00%	1/14/2019 22:51	100.00	ug/L
S190105-CCV6	Chromium KED ICPMS 200.8	304.19	ug/L	101.00%	1/14/2019 22:51	300.00	ug/L
S190105-CCV6	Copper KED ICPMS 200.8	103.40	ug/L	103.00%	1/14/2019 22:51	100.00	ug/L
S190105-CCV6	Lead KED ICPMS 200.8	48.67	ug/L	97.30%	1/14/2019 22:51	50.00	ug/L
S190105-CCV6	Nickel KED ICPMS 200.8	103.00	ug/L	103.00%	1/14/2019 22:51	100.00	ug/L
S190105-CCV6	Silver KED ICPMS 200.8	50.06	ug/L	100.00%	1/14/2019 22:51	50.00	ug/L
S190105-CCV6	Zinc KED ICPMS 200.8	208.34	ug/L	104.00%	1/14/2019 22:51	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190113-ICV1	Arsenic KED ICPMS 200.8	41.66	ug/L	104.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Arsenic KED ICPMS ELUT	41.66	ug/L	104.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Cadmium KED ICPMS 200.8	82.26	ug/L	103.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Cadmium KED ICPMS ELUT	82.26	ug/L	103.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Chromium KED ICPMS 200.8	246.60	ug/L	103.00%	1/31/2019 13:59	240.00	ug/L
S190113-ICV1	Chromium KED ICPMS ELUT	246.60	ug/L	103.00%	1/31/2019 13:59	240.00	ug/L
S190113-ICV1	Copper KED ICPMS 200.8	83.76	ug/L	105.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Copper KED ICPMS ELUT	83.76	ug/L	105.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Lead KED ICPMS 200.8	42.20	ug/L	106.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Lead KED ICPMS ELUT	42.20	ug/L	106.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Nickel KED ICPMS 200.8	83.41	ug/L	104.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Nickel KED ICPMS ELUT	83.41	ug/L	104.00%	1/31/2019 13:59	80.00	ug/L
S190113-ICV1	Silver KED ICPMS 200.8	41.25	ug/L	103.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Silver KED ICPMS ELUT	41.25	ug/L	103.00%	1/31/2019 13:59	40.00	ug/L
S190113-ICV1	Zinc KED ICPMS 200.8	170.99	ug/L	107.00%	1/31/2019 13:59	160.00	ug/L
S190113-ICV1	Zinc KED ICPMS ELUT	170.99	ug/L	107.00%	1/31/2019 13:59	160.00	ug/L
S190113-CCV1	Arsenic KED ICPMS 200.8	51.28	ug/L	103.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Arsenic KED ICPMS ELUT	51.28	ug/L	103.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Cadmium KED ICPMS 200.8	101.40	ug/L	101.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Cadmium KED ICPMS ELUT	101.40	ug/L	101.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Chromium KED ICPMS 200.8	305.13	ug/L	102.00%	1/31/2019 14:23	300.00	ug/L
S190113-CCV1	Chromium KED ICPMS ELUT	305.13	ug/L	102.00%	1/31/2019 14:23	300.00	ug/L
S190113-CCV1	Copper KED ICPMS 200.8	102.35	ug/L	102.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Copper KED ICPMS ELUT	102.35	ug/L	102.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Lead KED ICPMS 200.8	50.90	ug/L	102.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Lead KED ICPMS ELUT	50.90	ug/L	102.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Nickel KED ICPMS 200.8	102.66	ug/L	103.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Nickel KED ICPMS ELUT	102.66	ug/L	103.00%	1/31/2019 14:23	100.00	ug/L
S190113-CCV1	Silver KED ICPMS 200.8	50.58	ug/L	101.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Silver KED ICPMS ELUT	50.58	ug/L	101.00%	1/31/2019 14:23	50.00	ug/L
S190113-CCV1	Zinc KED ICPMS 200.8	208.67	ug/L	104.00%	1/31/2019 14:23	200.00	ug/L
S190113-CCV1	Zinc KED ICPMS ELUT	208.67	ug/L	104.00%	1/31/2019 14:23	200.00	ug/L
S190113-CCV2	Arsenic KED ICPMS 200.8	50.36	ug/L	101.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Arsenic KED ICPMS ELUT	50.36	ug/L	101.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Cadmium KED ICPMS 200.8	99.88	ug/L	99.90%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Cadmium KED ICPMS ELUT	99.88	ug/L	99.90%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Chromium KED ICPMS 200.8	300.72	ug/L	100.00%	1/31/2019 15:27	300.00	ug/L
S190113-CCV2	Chromium KED ICPMS ELUT	300.72	ug/L	100.00%	1/31/2019 15:27	300.00	ug/L
S190113-CCV2	Copper KED ICPMS 200.8	99.32	ug/L	99.30%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Copper KED ICPMS ELUT	99.32	ug/L	99.30%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Lead KED ICPMS 200.8	50.09	ug/L	100.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Lead KED ICPMS ELUT	50.09	ug/L	100.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Nickel KED ICPMS 200.8	100.25	ug/L	100.00%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Nickel KED ICPMS ELUT	100.25	ug/L	100.00%	1/31/2019 15:27	100.00	ug/L
S190113-CCV2	Silver KED ICPMS 200.8	49.51	ug/L	99.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Silver KED ICPMS ELUT	49.51	ug/L	99.00%	1/31/2019 15:27	50.00	ug/L
S190113-CCV2	Zinc KED ICPMS 200.8	204.87	ug/L	102.00%	1/31/2019 15:27	200.00	ug/L
S190113-CCV2	Zinc KED ICPMS ELUT	204.87	ug/L	102.00%	1/31/2019 15:27	200.00	ug/L
S190113-CCV3	Arsenic KED ICPMS 200.8	50.76	ug/L	102.00%	1/31/2019 16:37	50.00	ug/L
S190113-CCV3	Arsenic KED ICPMS ELUT	50.76	ug/L	102.00%	1/31/2019 16:37	50.00	ug/L

S190113-CCV3	Cadmium KED ICPMS 200.8	99.23	ug/L	99.20%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Cadmium KED ICPMS ELUT	99.23	ug/L	99.20%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Chromium KED ICPMS 200.8	301.17	ug/L	100.00%	1/31/2019 16:37	300.00	ug/L
S190113-CCV3	Chromium KED ICPMS ELUT	301.17	ug/L	100.00%	1/31/2019 16:37	300.00	ug/L
S190113-CCV3	Copper KED ICPMS 200.8	99.92	ug/L	99.90%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Copper KED ICPMS ELUT	99.92	ug/L	99.90%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Lead KED ICPMS 200.8	50.46	ug/L	101.00%	1/31/2019 16:37	50.00	ug/L
S190113-CCV3	Lead KED ICPMS ELUT	50.46	ug/L	101.00%	1/31/2019 16:37	50.00	ug/L
S190113-CCV3	Nickel KED ICPMS 200.8	100.12	ug/L	100.00%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Nickel KED ICPMS ELUT	100.12	ug/L	100.00%	1/31/2019 16:37	100.00	ug/L
S190113-CCV3	Silver KED ICPMS 200.8	48.54	ug/L	97.10%	1/31/2019 16:37	50.00	ug/L
S190113-CCV3	Silver KED ICPMS ELUT	48.54	ug/L	97.10%	1/31/2019 16:37	50.00	ug/L
S190113-CCV3	Zinc KED ICPMS 200.8	205.80	ug/L	103.00%	1/31/2019 16:37	200.00	ug/L
S190113-CCV3	Zinc KED ICPMS ELUT	205.80	ug/L	103.00%	1/31/2019 16:37	200.00	ug/L
S190113-CCV4	Arsenic KED ICPMS 200.8	50.40	ug/L	101.00%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Arsenic KED ICPMS ELUT	50.40	ug/L	101.00%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Cadmium KED ICPMS 200.8	98.35	ug/L	98.40%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Cadmium KED ICPMS ELUT	98.35	ug/L	98.40%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Chromium KED ICPMS 200.8	291.04	ug/L	97.00%	1/31/2019 17:47	300.00	ug/L
S190113-CCV4	Chromium KED ICPMS ELUT	291.04	ug/L	97.00%	1/31/2019 17:47	300.00	ug/L
S190113-CCV4	Copper KED ICPMS 200.8	98.29	ug/L	98.30%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Copper KED ICPMS ELUT	98.29	ug/L	98.30%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Lead KED ICPMS 200.8	51.30	ug/L	103.00%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Lead KED ICPMS ELUT	51.30	ug/L	103.00%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Nickel KED ICPMS 200.8	97.24	ug/L	97.20%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Nickel KED ICPMS ELUT	97.24	ug/L	97.20%	1/31/2019 17:47	100.00	ug/L
S190113-CCV4	Silver KED ICPMS 200.8	47.86	ug/L	95.70%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Silver KED ICPMS ELUT	47.86	ug/L	95.70%	1/31/2019 17:47	50.00	ug/L
S190113-CCV4	Zinc KED ICPMS 200.8	204.16	ug/L	102.00%	1/31/2019 17:47	200.00	ug/L
S190113-CCV4	Zinc KED ICPMS ELUT	204.16	ug/L	102.00%	1/31/2019 17:47	200.00	ug/L
S190113-CCV5	Arsenic KED ICPMS 200.8	51.07	ug/L	102.00%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Arsenic KED ICPMS ELUT	51.07	ug/L	102.00%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Cadmium KED ICPMS 200.8	98.97	ug/L	99.00%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Cadmium KED ICPMS ELUT	98.97	ug/L	99.00%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Chromium KED ICPMS 200.8	295.34	ug/L	98.40%	1/31/2019 18:56	300.00	ug/L
S190113-CCV5	Chromium KED ICPMS ELUT	295.34	ug/L	98.40%	1/31/2019 18:56	300.00	ug/L
S190113-CCV5	Copper KED ICPMS 200.8	98.59	ug/L	98.60%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Copper KED ICPMS ELUT	98.59	ug/L	98.60%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Lead KED ICPMS 200.8	50.88	ug/L	102.00%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Lead KED ICPMS ELUT	50.88	ug/L	102.00%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Nickel KED ICPMS 200.8	97.96	ug/L	98.00%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Nickel KED ICPMS ELUT	97.96	ug/L	98.00%	1/31/2019 18:56	100.00	ug/L
S190113-CCV5	Silver KED ICPMS 200.8	47.80	ug/L	95.60%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Silver KED ICPMS ELUT	47.80	ug/L	95.60%	1/31/2019 18:56	50.00	ug/L
S190113-CCV5	Zinc KED ICPMS 200.8	205.62	ug/L	103.00%	1/31/2019 18:56	200.00	ug/L
S190113-CCV5	Zinc KED ICPMS ELUT	205.62	ug/L	103.00%	1/31/2019 18:56	200.00	ug/L
S190113-CCV6	Arsenic KED ICPMS 200.8	50.18	ug/L	100.00%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Arsenic KED ICPMS ELUT	50.18	ug/L	100.00%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Cadmium KED ICPMS 200.8	98.06	ug/L	98.10%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Cadmium KED ICPMS ELUT	98.06	ug/L	98.10%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Chromium KED ICPMS 200.8	286.04	ug/L	95.30%	1/31/2019 20:06	300.00	ug/L
S190113-CCV6	Chromium KED ICPMS ELUT	286.04	ug/L	95.30%	1/31/2019 20:06	300.00	ug/L

S190113-CCV6	Copper KED ICPMS 200.8	97.07	ug/L	97.10%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Copper KED ICPMS ELUT	97.07	ug/L	97.10%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Lead KED ICPMS 200.8	51.52	ug/L	103.00%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Lead KED ICPMS ELUT	51.52	ug/L	103.00%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Nickel KED ICPMS 200.8	95.48	ug/L	95.50%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Nickel KED ICPMS ELUT	95.48	ug/L	95.50%	1/31/2019 20:06	100.00	ug/L
S190113-CCV6	Silver KED ICPMS 200.8	46.96	ug/L	93.90%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Silver KED ICPMS ELUT	46.96	ug/L	93.90%	1/31/2019 20:06	50.00	ug/L
S190113-CCV6	Zinc KED ICPMS 200.8	204.21	ug/L	102.00%	1/31/2019 20:06	200.00	ug/L
S190113-CCV6	Zinc KED ICPMS ELUT	204.21	ug/L	102.00%	1/31/2019 20:06	200.00	ug/L
S190113-CCV7	Arsenic KED ICPMS 200.8	50.63	ug/L	101.00%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Arsenic KED ICPMS ELUT	50.63	ug/L	101.00%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Cadmium KED ICPMS 200.8	98.08	ug/L	98.10%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Cadmium KED ICPMS ELUT	98.08	ug/L	98.10%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Chromium KED ICPMS 200.8	287.27	ug/L	95.80%	1/31/2019 20:49	300.00	ug/L
S190113-CCV7	Chromium KED ICPMS ELUT	287.27	ug/L	95.80%	1/31/2019 20:49	300.00	ug/L
S190113-CCV7	Copper KED ICPMS 200.8	96.40	ug/L	96.40%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Copper KED ICPMS ELUT	96.40	ug/L	96.40%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Lead KED ICPMS 200.8	51.48	ug/L	103.00%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Lead KED ICPMS ELUT	51.48	ug/L	103.00%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Nickel KED ICPMS 200.8	95.35	ug/L	95.30%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Nickel KED ICPMS ELUT	95.35	ug/L	95.30%	1/31/2019 20:49	100.00	ug/L
S190113-CCV7	Silver KED ICPMS 200.8	46.80	ug/L	93.60%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Silver KED ICPMS ELUT	46.80	ug/L	93.60%	1/31/2019 20:49	50.00	ug/L
S190113-CCV7	Zinc KED ICPMS 200.8	203.63	ug/L	102.00%	1/31/2019 20:49	200.00	ug/L
S190113-CCV7	Zinc KED ICPMS ELUT	203.63	ug/L	102.00%	1/31/2019 20:49	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190202-ICV1	Arsenic KED ICPMS 200.8	42.20	ug/L	105.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Arsenic KED ICPMS ELUT	42.20	ug/L	105.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Cadmium KED ICPMS 200.8	81.86	ug/L	102.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Cadmium KED ICPMS ELUT	81.86	ug/L	102.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Chromium KED ICPMS 200.8	247.82	ug/L	103.00%	2/1/2019 15:32	240.00	ug/L
S190202-ICV1	Chromium KED ICPMS ELUT	247.82	ug/L	103.00%	2/1/2019 15:32	240.00	ug/L
S190202-ICV1	Copper KED ICPMS 200.8	84.34	ug/L	105.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Copper KED ICPMS ELUT	84.34	ug/L	105.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Lead KED ICPMS 200.8	42.42	ug/L	106.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Lead KED ICPMS ELUT	42.42	ug/L	106.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Nickel KED ICPMS 200.8	84.07	ug/L	105.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Nickel KED ICPMS ELUT	84.07	ug/L	105.00%	2/1/2019 15:32	80.00	ug/L
S190202-ICV1	Silver KED ICPMS 200.8	40.49	ug/L	101.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Silver KED ICPMS ELUT	40.49	ug/L	101.00%	2/1/2019 15:32	40.00	ug/L
S190202-ICV1	Zinc KED ICPMS 200.8	172.29	ug/L	108.00%	2/1/2019 15:32	160.00	ug/L
S190202-ICV1	Zinc KED ICPMS ELUT	172.29	ug/L	108.00%	2/1/2019 15:32	160.00	ug/L
S190202-CCV1	Arsenic KED ICPMS 200.8	51.29	ug/L	103.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Arsenic KED ICPMS ELUT	51.29	ug/L	103.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Cadmium KED ICPMS 200.8	100.92	ug/L	101.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Cadmium KED ICPMS ELUT	100.92	ug/L	101.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Chromium KED ICPMS 200.8	305.75	ug/L	102.00%	2/1/2019 15:57	300.00	ug/L
S190202-CCV1	Chromium KED ICPMS ELUT	305.75	ug/L	102.00%	2/1/2019 15:57	300.00	ug/L
S190202-CCV1	Copper KED ICPMS 200.8	102.52	ug/L	103.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Copper KED ICPMS ELUT	102.52	ug/L	103.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Lead KED ICPMS 200.8	51.09	ug/L	102.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Lead KED ICPMS ELUT	51.09	ug/L	102.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Nickel KED ICPMS 200.8	102.88	ug/L	103.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Nickel KED ICPMS ELUT	102.88	ug/L	103.00%	2/1/2019 15:57	100.00	ug/L
S190202-CCV1	Silver KED ICPMS 200.8	50.10	ug/L	100.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Silver KED ICPMS ELUT	50.10	ug/L	100.00%	2/1/2019 15:57	50.00	ug/L
S190202-CCV1	Zinc KED ICPMS 200.8	210.85	ug/L	105.00%	2/1/2019 15:57	200.00	ug/L
S190202-CCV1	Zinc KED ICPMS ELUT	210.85	ug/L	105.00%	2/1/2019 15:57	200.00	ug/L
S190202-CCV2	Arsenic KED ICPMS 200.8	51.00	ug/L	102.00%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Arsenic KED ICPMS ELUT	51.00	ug/L	102.00%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Cadmium KED ICPMS 200.8	100.05	ug/L	100.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Cadmium KED ICPMS ELUT	100.05	ug/L	100.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Chromium KED ICPMS 200.8	302.68	ug/L	101.00%	2/1/2019 17:01	300.00	ug/L
S190202-CCV2	Chromium KED ICPMS ELUT	302.68	ug/L	101.00%	2/1/2019 17:01	300.00	ug/L
S190202-CCV2	Copper KED ICPMS 200.8	100.56	ug/L	101.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Copper KED ICPMS ELUT	100.56	ug/L	101.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Lead KED ICPMS 200.8	51.02	ug/L	102.00%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Lead KED ICPMS ELUT	51.02	ug/L	102.00%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Nickel KED ICPMS 200.8	100.68	ug/L	101.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Nickel KED ICPMS ELUT	100.68	ug/L	101.00%	2/1/2019 17:01	100.00	ug/L
S190202-CCV2	Silver KED ICPMS 200.8	49.30	ug/L	98.60%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Silver KED ICPMS ELUT	49.30	ug/L	98.60%	2/1/2019 17:01	50.00	ug/L
S190202-CCV2	Zinc KED ICPMS 200.8	207.72	ug/L	104.00%	2/1/2019 17:01	200.00	ug/L
S190202-CCV2	Zinc KED ICPMS ELUT	207.72	ug/L	104.00%	2/1/2019 17:01	200.00	ug/L
S190202-CCV3	Arsenic KED ICPMS 200.8	50.88	ug/L	102.00%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Arsenic KED ICPMS ELUT	50.88	ug/L	102.00%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Cadmium KED ICPMS 200.8	99.63	ug/L	99.60%	2/1/2019 18:10	100.00	ug/L

S190202-CCV3	Cadmium KED ICPMS ELUT	99.63	ug/L	99.60%	2/1/2019 18:10	100.00	ug/L
S190202-CCV3	Chromium KED ICPMS 200.8	303.30	ug/L	101.00%	2/1/2019 18:10	300.00	ug/L
S190202-CCV3	Chromium KED ICPMS ELUT	303.30	ug/L	101.00%	2/1/2019 18:10	300.00	ug/L
S190202-CCV3	Copper KED ICPMS 200.8	101.35	ug/L	101.00%	2/1/2019 18:10	100.00	ug/L
S190202-CCV3	Copper KED ICPMS ELUT	101.35	ug/L	101.00%	2/1/2019 18:10	100.00	ug/L
S190202-CCV3	Lead KED ICPMS 200.8	50.63	ug/L	101.00%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Lead KED ICPMS ELUT	50.63	ug/L	101.00%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Nickel KED ICPMS 200.8	101.26	ug/L	101.00%	2/1/2019 18:10	100.00	ug/L
S190202-CCV3	Nickel KED ICPMS ELUT	101.26	ug/L	101.00%	2/1/2019 18:10	100.00	ug/L
S190202-CCV3	Silver KED ICPMS 200.8	49.41	ug/L	98.80%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Silver KED ICPMS ELUT	49.41	ug/L	98.80%	2/1/2019 18:10	50.00	ug/L
S190202-CCV3	Zinc KED ICPMS 200.8	208.97	ug/L	104.00%	2/1/2019 18:10	200.00	ug/L
S190202-CCV3	Zinc KED ICPMS ELUT	208.97	ug/L	104.00%	2/1/2019 18:10	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190203-ICV1	Arsenic KED ICPMS 200.8	40.65	ug/L	102.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Arsenic KED ICPMS ELUT	40.65	ug/L	102.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Cadmium KED ICPMS 200.8	81.26	ug/L	102.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Cadmium KED ICPMS ELUT	81.26	ug/L	102.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Chromium KED ICPMS 200.8	242.63	ug/L	101.00%	2/4/2019 15:45	240.00	ug/L
S190203-ICV1	Chromium KED ICPMS ELUT	242.63	ug/L	101.00%	2/4/2019 15:45	240.00	ug/L
S190203-ICV1	Copper KED ICPMS 200.8	82.42	ug/L	103.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Copper KED ICPMS ELUT	82.42	ug/L	103.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Lead KED ICPMS 200.8	41.50	ug/L	104.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Lead KED ICPMS ELUT	41.50	ug/L	104.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Nickel KED ICPMS 200.8	82.34	ug/L	103.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Nickel KED ICPMS ELUT	82.34	ug/L	103.00%	2/4/2019 15:45	80.00	ug/L
S190203-ICV1	Silver KED ICPMS 200.8	40.73	ug/L	102.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Silver KED ICPMS ELUT	40.73	ug/L	102.00%	2/4/2019 15:45	40.00	ug/L
S190203-ICV1	Zinc KED ICPMS 200.8	168.65	ug/L	105.00%	2/4/2019 15:45	160.00	ug/L
S190203-ICV1	Zinc KED ICPMS ELUT	168.65	ug/L	105.00%	2/4/2019 15:45	160.00	ug/L
S190203-CCV1	Arsenic KED ICPMS 200.8	50.49	ug/L	101.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Arsenic KED ICPMS ELUT	50.49	ug/L	101.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Cadmium KED ICPMS 200.8	101.11	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Cadmium KED ICPMS ELUT	101.11	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Chromium KED ICPMS 200.8	304.04	ug/L	101.00%	2/4/2019 16:09	300.00	ug/L
S190203-CCV1	Chromium KED ICPMS ELUT	304.04	ug/L	101.00%	2/4/2019 16:09	300.00	ug/L
S190203-CCV1	Copper KED ICPMS 200.8	101.18	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Copper KED ICPMS ELUT	101.18	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Lead KED ICPMS 200.8	51.12	ug/L	102.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Lead KED ICPMS ELUT	51.12	ug/L	102.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Nickel KED ICPMS 200.8	101.11	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Nickel KED ICPMS ELUT	101.11	ug/L	101.00%	2/4/2019 16:09	100.00	ug/L
S190203-CCV1	Silver KED ICPMS 200.8	50.19	ug/L	100.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Silver KED ICPMS ELUT	50.19	ug/L	100.00%	2/4/2019 16:09	50.00	ug/L
S190203-CCV1	Zinc KED ICPMS 200.8	206.85	ug/L	103.00%	2/4/2019 16:09	200.00	ug/L
S190203-CCV1	Zinc KED ICPMS ELUT	206.85	ug/L	103.00%	2/4/2019 16:09	200.00	ug/L
S190203-CCV2	Arsenic KED ICPMS 200.8	50.39	ug/L	101.00%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Arsenic KED ICPMS ELUT	50.39	ug/L	101.00%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Cadmium KED ICPMS 200.8	98.68	ug/L	98.70%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Cadmium KED ICPMS ELUT	98.68	ug/L	98.70%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Chromium KED ICPMS 200.8	288.33	ug/L	96.10%	2/4/2019 17:03	300.00	ug/L
S190203-CCV2	Chromium KED ICPMS ELUT	288.33	ug/L	96.10%	2/4/2019 17:03	300.00	ug/L
S190203-CCV2	Copper KED ICPMS 200.8	96.26	ug/L	96.30%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Copper KED ICPMS ELUT	96.26	ug/L	96.30%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Lead KED ICPMS 200.8	49.60	ug/L	99.20%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Lead KED ICPMS ELUT	49.60	ug/L	99.20%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Nickel KED ICPMS 200.8	96.75	ug/L	96.80%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Nickel KED ICPMS ELUT	96.75	ug/L	96.80%	2/4/2019 17:03	100.00	ug/L
S190203-CCV2	Silver KED ICPMS 200.8	48.28	ug/L	96.60%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Silver KED ICPMS ELUT	48.28	ug/L	96.60%	2/4/2019 17:03	50.00	ug/L
S190203-CCV2	Zinc KED ICPMS 200.8	201.05	ug/L	101.00%	2/4/2019 17:03	200.00	ug/L

S190203-CCV2	Zinc KED ICPMS ELUT	201.05	ug/L	101.00%	2/4/2019 17:03	200.00	ug/L
S190203-CCV3	Arsenic KED ICPMS 200.8	50.34	ug/L	101.00%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Arsenic KED ICPMS ELUT	50.34	ug/L	101.00%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Cadmium KED ICPMS 200.8	99.36	ug/L	99.40%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Cadmium KED ICPMS ELUT	99.36	ug/L	99.40%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Chromium KED ICPMS 200.8	291.76	ug/L	97.30%	2/4/2019 18:13	300.00	ug/L
S190203-CCV3	Chromium KED ICPMS ELUT	291.76	ug/L	97.30%	2/4/2019 18:13	300.00	ug/L
S190203-CCV3	Copper KED ICPMS 200.8	96.84	ug/L	96.80%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Copper KED ICPMS ELUT	96.84	ug/L	96.80%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Lead KED ICPMS 200.8	50.65	ug/L	101.00%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Lead KED ICPMS ELUT	50.65	ug/L	101.00%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Nickel KED ICPMS 200.8	97.25	ug/L	97.30%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Nickel KED ICPMS ELUT	97.25	ug/L	97.30%	2/4/2019 18:13	100.00	ug/L
S190203-CCV3	Silver KED ICPMS 200.8	48.53	ug/L	97.10%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Silver KED ICPMS ELUT	48.53	ug/L	97.10%	2/4/2019 18:13	50.00	ug/L
S190203-CCV3	Zinc KED ICPMS 200.8	201.96	ug/L	101.00%	2/4/2019 18:13	200.00	ug/L
S190203-CCV3	Zinc KED ICPMS ELUT	201.96	ug/L	101.00%	2/4/2019 18:13	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190204-ICV1	Arsenic KED ICPMS 200.8	39.89	ug/L	99.70%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Arsenic KED ICPMS ELUT	39.89	ug/L	99.70%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Cadmium KED ICPMS 200.8	80.02	ug/L	100.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Cadmium KED ICPMS ELUT	80.02	ug/L	100.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Chromium KED ICPMS 200.8	238.46	ug/L	99.40%	2/5/2019 15:16	240.00	ug/L
S190204-ICV1	Chromium KED ICPMS ELUT	238.46	ug/L	99.40%	2/5/2019 15:16	240.00	ug/L
S190204-ICV1	Copper KED ICPMS 200.8	80.84	ug/L	101.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Copper KED ICPMS ELUT	80.84	ug/L	101.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Lead KED ICPMS 200.8	41.30	ug/L	103.00%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Lead KED ICPMS ELUT	41.30	ug/L	103.00%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Nickel KED ICPMS 200.8	80.94	ug/L	101.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Nickel KED ICPMS ELUT	80.94	ug/L	101.00%	2/5/2019 15:16	80.00	ug/L
S190204-ICV1	Silver KED ICPMS 200.8	40.37	ug/L	101.00%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Silver KED ICPMS ELUT	40.37	ug/L	101.00%	2/5/2019 15:16	40.00	ug/L
S190204-ICV1	Zinc KED ICPMS 200.8	166.14	ug/L	104.00%	2/5/2019 15:16	160.00	ug/L
S190204-ICV1	Zinc KED ICPMS ELUT	166.14	ug/L	104.00%	2/5/2019 15:16	160.00	ug/L
S190204-CCV1	Arsenic KED ICPMS 200.8	50.75	ug/L	102.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Arsenic KED ICPMS ELUT	50.75	ug/L	102.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Cadmium KED ICPMS 200.8	101.30	ug/L	101.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Cadmium KED ICPMS ELUT	101.30	ug/L	101.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Chromium KED ICPMS 200.8	302.14	ug/L	101.00%	2/5/2019 15:41	300.00	ug/L
S190204-CCV1	Chromium KED ICPMS ELUT	302.14	ug/L	101.00%	2/5/2019 15:41	300.00	ug/L
S190204-CCV1	Copper KED ICPMS 200.8	101.57	ug/L	102.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Copper KED ICPMS ELUT	101.57	ug/L	102.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Lead KED ICPMS 200.8	50.60	ug/L	101.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Lead KED ICPMS ELUT	50.60	ug/L	101.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Nickel KED ICPMS 200.8	101.36	ug/L	101.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Nickel KED ICPMS ELUT	101.36	ug/L	101.00%	2/5/2019 15:41	100.00	ug/L
S190204-CCV1	Silver KED ICPMS 200.8	50.53	ug/L	101.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Silver KED ICPMS ELUT	50.53	ug/L	101.00%	2/5/2019 15:41	50.00	ug/L
S190204-CCV1	Zinc KED ICPMS 200.8	207.35	ug/L	104.00%	2/5/2019 15:41	200.00	ug/L
S190204-CCV1	Zinc KED ICPMS ELUT	207.35	ug/L	104.00%	2/5/2019 15:41	200.00	ug/L
S190204-CCV2	Arsenic KED ICPMS 200.8	50.75	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Arsenic KED ICPMS ELUT	50.75	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Cadmium KED ICPMS 200.8	100.42	ug/L	100.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Cadmium KED ICPMS ELUT	100.42	ug/L	100.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Chromium KED ICPMS 200.8	304.87	ug/L	102.00%	2/5/2019 16:39	300.00	ug/L
S190204-CCV2	Chromium KED ICPMS ELUT	304.87	ug/L	102.00%	2/5/2019 16:39	300.00	ug/L
S190204-CCV2	Copper KED ICPMS 200.8	101.55	ug/L	102.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Copper KED ICPMS ELUT	101.55	ug/L	102.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Lead KED ICPMS 200.8	50.73	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Lead KED ICPMS ELUT	50.73	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Nickel KED ICPMS 200.8	102.60	ug/L	103.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Nickel KED ICPMS ELUT	102.60	ug/L	103.00%	2/5/2019 16:39	100.00	ug/L
S190204-CCV2	Silver KED ICPMS 200.8	50.28	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Silver KED ICPMS ELUT	50.28	ug/L	101.00%	2/5/2019 16:39	50.00	ug/L
S190204-CCV2	Zinc KED ICPMS 200.8	206.82	ug/L	103.00%	2/5/2019 16:39	200.00	ug/L
S190204-CCV2	Zinc KED ICPMS ELUT	206.82	ug/L	103.00%	2/5/2019 16:39	200.00	ug/L

S190204-CCV3	Arsenic KED ICPMS 200.8	50.66	ug/L	101.00%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Arsenic KED ICPMS ELUT	50.66	ug/L	101.00%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Cadmium KED ICPMS 200.8	101.41	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Cadmium KED ICPMS ELUT	101.41	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Chromium KED ICPMS 200.8	304.40	ug/L	101.00%	2/5/2019 17:49	300.00	ug/L
S190204-CCV3	Chromium KED ICPMS ELUT	304.40	ug/L	101.00%	2/5/2019 17:49	300.00	ug/L
S190204-CCV3	Copper KED ICPMS 200.8	101.03	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Copper KED ICPMS ELUT	101.03	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Lead KED ICPMS 200.8	50.04	ug/L	100.00%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Lead KED ICPMS ELUT	50.04	ug/L	100.00%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Nickel KED ICPMS 200.8	101.16	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Nickel KED ICPMS ELUT	101.16	ug/L	101.00%	2/5/2019 17:49	100.00	ug/L
S190204-CCV3	Silver KED ICPMS 200.8	49.75	ug/L	99.50%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Silver KED ICPMS ELUT	49.75	ug/L	99.50%	2/5/2019 17:49	50.00	ug/L
S190204-CCV3	Zinc KED ICPMS 200.8	208.88	ug/L	104.00%	2/5/2019 17:49	200.00	ug/L
S190204-CCV3	Zinc KED ICPMS ELUT	208.88	ug/L	104.00%	2/5/2019 17:49	200.00	ug/L
S190204-CCV4	Arsenic KED ICPMS 200.8	49.62	ug/L	99.20%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Arsenic KED ICPMS ELUT	49.62	ug/L	99.20%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Cadmium KED ICPMS 200.8	99.32	ug/L	99.30%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Cadmium KED ICPMS ELUT	99.32	ug/L	99.30%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Chromium KED ICPMS 200.8	292.15	ug/L	97.40%	2/5/2019 19:04	300.00	ug/L
S190204-CCV4	Chromium KED ICPMS ELUT	292.15	ug/L	97.40%	2/5/2019 19:04	300.00	ug/L
S190204-CCV4	Copper KED ICPMS 200.8	96.57	ug/L	96.60%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Copper KED ICPMS ELUT	96.57	ug/L	96.60%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Lead KED ICPMS 200.8	47.85	ug/L	95.70%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Lead KED ICPMS ELUT	47.85	ug/L	95.70%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Nickel KED ICPMS 200.8	96.48	ug/L	96.50%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Nickel KED ICPMS ELUT	96.48	ug/L	96.50%	2/5/2019 19:04	100.00	ug/L
S190204-CCV4	Silver KED ICPMS 200.8	48.70	ug/L	97.40%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Silver KED ICPMS ELUT	48.70	ug/L	97.40%	2/5/2019 19:04	50.00	ug/L
S190204-CCV4	Zinc KED ICPMS 200.8	203.94	ug/L	102.00%	2/5/2019 19:04	200.00	ug/L
S190204-CCV4	Zinc KED ICPMS ELUT	203.94	ug/L	102.00%	2/5/2019 19:04	200.00	ug/L
S190204-CCV5	Arsenic KED ICPMS 200.8	49.78	ug/L	99.60%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Arsenic KED ICPMS ELUT	49.78	ug/L	99.60%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Cadmium KED ICPMS 200.8	99.48	ug/L	99.50%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Cadmium KED ICPMS ELUT	99.48	ug/L	99.50%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Chromium KED ICPMS 200.8	291.62	ug/L	97.20%	2/5/2019 20:13	300.00	ug/L
S190204-CCV5	Chromium KED ICPMS ELUT	291.62	ug/L	97.20%	2/5/2019 20:13	300.00	ug/L
S190204-CCV5	Copper KED ICPMS 200.8	98.11	ug/L	98.10%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Copper KED ICPMS ELUT	98.11	ug/L	98.10%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Lead KED ICPMS 200.8	47.98	ug/L	96.00%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Lead KED ICPMS ELUT	47.98	ug/L	96.00%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Nickel KED ICPMS 200.8	96.73	ug/L	96.70%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Nickel KED ICPMS ELUT	96.73	ug/L	96.70%	2/5/2019 20:13	100.00	ug/L
S190204-CCV5	Silver KED ICPMS 200.8	48.66	ug/L	97.30%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Silver KED ICPMS ELUT	48.66	ug/L	97.30%	2/5/2019 20:13	50.00	ug/L
S190204-CCV5	Zinc KED ICPMS 200.8	205.14	ug/L	103.00%	2/5/2019 20:13	200.00	ug/L
S190204-CCV5	Zinc KED ICPMS ELUT	205.14	ug/L	103.00%	2/5/2019 20:13	200.00	ug/L
S190204-CCV6	Arsenic KED ICPMS 200.8	49.12	ug/L	98.20%	2/5/2019 21:23	50.00	ug/L
S190204-CCV6	Arsenic KED ICPMS ELUT	49.12	ug/L	98.20%	2/5/2019 21:23	50.00	ug/L

S190204-CCV6	Cadmium KED ICPMS 200.8	99.82	ug/L	99.80%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Cadmium KED ICPMS ELUT	99.82	ug/L	99.80%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Chromium KED ICPMS 200.8	287.44	ug/L	95.80%	2/5/2019 21:23	300.00	ug/L
S190204-CCV6	Chromium KED ICPMS ELUT	287.44	ug/L	95.80%	2/5/2019 21:23	300.00	ug/L
S190204-CCV6	Copper KED ICPMS 200.8	96.00	ug/L	96.00%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Copper KED ICPMS ELUT	96.00	ug/L	96.00%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Lead KED ICPMS 200.8	47.41	ug/L	94.80%	2/5/2019 21:23	50.00	ug/L
S190204-CCV6	Lead KED ICPMS ELUT	47.41	ug/L	94.80%	2/5/2019 21:23	50.00	ug/L
S190204-CCV6	Nickel KED ICPMS 200.8	95.60	ug/L	95.60%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Nickel KED ICPMS ELUT	95.60	ug/L	95.60%	2/5/2019 21:23	100.00	ug/L
S190204-CCV6	Silver KED ICPMS 200.8	49.09	ug/L	98.20%	2/5/2019 21:23	50.00	ug/L
S190204-CCV6	Silver KED ICPMS ELUT	49.09	ug/L	98.20%	2/5/2019 21:23	50.00	ug/L
S190204-CCV6	Zinc KED ICPMS 200.8	201.27	ug/L	101.00%	2/5/2019 21:23	200.00	ug/L
S190204-CCV6	Zinc KED ICPMS ELUT	201.27	ug/L	101.00%	2/5/2019 21:23	200.00	ug/L
S190204-CCV7	Arsenic KED ICPMS 200.8	49.58	ug/L	99.20%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Arsenic KED ICPMS ELUT	49.58	ug/L	99.20%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Cadmium KED ICPMS 200.8	100.49	ug/L	100.00%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Cadmium KED ICPMS ELUT	100.49	ug/L	100.00%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Chromium KED ICPMS 200.8	288.84	ug/L	96.30%	2/5/2019 22:33	300.00	ug/L
S190204-CCV7	Chromium KED ICPMS ELUT	288.84	ug/L	96.30%	2/5/2019 22:33	300.00	ug/L
S190204-CCV7	Copper KED ICPMS 200.8	96.43	ug/L	96.40%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Copper KED ICPMS ELUT	96.43	ug/L	96.40%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Lead KED ICPMS 200.8	47.15	ug/L	94.30%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Lead KED ICPMS ELUT	47.15	ug/L	94.30%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Nickel KED ICPMS 200.8	95.71	ug/L	95.70%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Nickel KED ICPMS ELUT	95.71	ug/L	95.70%	2/5/2019 22:33	100.00	ug/L
S190204-CCV7	Silver KED ICPMS 200.8	49.21	ug/L	98.40%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Silver KED ICPMS ELUT	49.21	ug/L	98.40%	2/5/2019 22:33	50.00	ug/L
S190204-CCV7	Zinc KED ICPMS 200.8	203.30	ug/L	102.00%	2/5/2019 22:33	200.00	ug/L
S190204-CCV7	Zinc KED ICPMS ELUT	203.30	ug/L	102.00%	2/5/2019 22:33	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190205-ICV1	Arsenic KED ICPMS 200.8	40.41	ug/L	101.00%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Arsenic KED ICPMS ELUT	40.41	ug/L	101.00%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Cadmium KED ICPMS 200.8	79.72	ug/L	99.70%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Cadmium KED ICPMS ELUT	79.72	ug/L	99.70%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Chromium KED ICPMS 200.8	239.70	ug/L	99.90%	2/6/2019 12:01	240.00	ug/L
S190205-ICV1	Chromium KED ICPMS ELUT	239.70	ug/L	99.90%	2/6/2019 12:01	240.00	ug/L
S190205-ICV1	Copper KED ICPMS 200.8	81.51	ug/L	102.00%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Copper KED ICPMS ELUT	81.51	ug/L	102.00%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Lead KED ICPMS 200.8	40.88	ug/L	102.00%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Lead KED ICPMS ELUT	40.88	ug/L	102.00%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Nickel KED ICPMS 200.8	80.86	ug/L	101.00%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Nickel KED ICPMS ELUT	80.86	ug/L	101.00%	2/6/2019 12:01	80.00	ug/L
S190205-ICV1	Silver KED ICPMS 200.8	39.71	ug/L	99.30%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Silver KED ICPMS ELUT	39.71	ug/L	99.30%	2/6/2019 12:01	40.00	ug/L
S190205-ICV1	Zinc KED ICPMS 200.8	167.15	ug/L	104.00%	2/6/2019 12:01	160.00	ug/L
S190205-ICV1	Zinc KED ICPMS ELUT	167.15	ug/L	104.00%	2/6/2019 12:01	160.00	ug/L
S190205-CCV1	Arsenic KED ICPMS 200.8	51.02	ug/L	102.00%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Arsenic KED ICPMS ELUT	51.02	ug/L	102.00%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Cadmium KED ICPMS 200.8	100.45	ug/L	100.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Cadmium KED ICPMS ELUT	100.45	ug/L	100.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Chromium KED ICPMS 200.8	304.94	ug/L	102.00%	2/6/2019 12:26	300.00	ug/L
S190205-CCV1	Chromium KED ICPMS ELUT	304.94	ug/L	102.00%	2/6/2019 12:26	300.00	ug/L
S190205-CCV1	Copper KED ICPMS 200.8	102.14	ug/L	102.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Copper KED ICPMS ELUT	102.14	ug/L	102.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Lead KED ICPMS 200.8	50.74	ug/L	101.00%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Lead KED ICPMS ELUT	50.74	ug/L	101.00%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Nickel KED ICPMS 200.8	101.21	ug/L	101.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Nickel KED ICPMS ELUT	101.21	ug/L	101.00%	2/6/2019 12:26	100.00	ug/L
S190205-CCV1	Silver KED ICPMS 200.8	49.88	ug/L	99.80%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Silver KED ICPMS ELUT	49.88	ug/L	99.80%	2/6/2019 12:26	50.00	ug/L
S190205-CCV1	Zinc KED ICPMS 200.8	210.28	ug/L	105.00%	2/6/2019 12:26	200.00	ug/L
S190205-CCV1	Zinc KED ICPMS ELUT	210.28	ug/L	105.00%	2/6/2019 12:26	200.00	ug/L
S190205-CCV2	Arsenic KED ICPMS 200.8	51.53	ug/L	103.00%	2/6/2019 13:02	50.00	ug/L
S190205-CCV2	Arsenic KED ICPMS ELUT	51.53	ug/L	103.00%	2/6/2019 13:02	50.00	ug/L
S190205-CCV2	Cadmium KED ICPMS 200.8	99.39	ug/L	99.40%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Cadmium KED ICPMS ELUT	99.39	ug/L	99.40%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Chromium KED ICPMS 200.8	301.61	ug/L	101.00%	2/6/2019 13:02	300.00	ug/L
S190205-CCV2	Chromium KED ICPMS ELUT	301.61	ug/L	101.00%	2/6/2019 13:02	300.00	ug/L
S190205-CCV2	Copper KED ICPMS 200.8	101.66	ug/L	102.00%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Copper KED ICPMS ELUT	101.66	ug/L	102.00%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Lead KED ICPMS 200.8	50.37	ug/L	101.00%	2/6/2019 13:02	50.00	ug/L
S190205-CCV2	Lead KED ICPMS ELUT	50.37	ug/L	101.00%	2/6/2019 13:02	50.00	ug/L
S190205-CCV2	Nickel KED ICPMS 200.8	100.31	ug/L	100.00%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Nickel KED ICPMS ELUT	100.31	ug/L	100.00%	2/6/2019 13:02	100.00	ug/L
S190205-CCV2	Silver KED ICPMS 200.8	49.43	ug/L	98.90%	2/6/2019 13:02	50.00	ug/L
S190205-CCV2	Silver KED ICPMS ELUT	49.43	ug/L	98.90%	2/6/2019 13:02	50.00	ug/L

S190205-CCV2	Zinc KED ICPMS 200.8	209.82	ug/L	105.00%	2/6/2019 13:02	200.00	ug/L
S190205-CCV2	Zinc KED ICPMS ELUT	209.82	ug/L	105.00%	2/6/2019 13:02	200.00	ug/L

LabNumber	Analysis	Units	Result	Recovery	Analyzed	Spike Conc	Spike Unit
S190319-ICV1	Arsenic KED ICPMS 200.8	ug/L	40.93	102.00%	3/12/2019 17:23	40.00	ug/L
S190319-ICV1	Cadmium KED ICPMS 200.8	ug/L	80.51	101.00%	3/12/2019 17:23	80.00	ug/L
S190319-ICV1	Chromium KED ICPMS 200.8	ug/L	242.94	101.00%	3/12/2019 17:23	240.00	ug/L
S190319-ICV1	Copper KED ICPMS 200.8	ug/L	81.66	102.00%	3/12/2019 17:23	80.00	ug/L
S190319-ICV1	Lead KED ICPMS 200.8	ug/L	40.78	102.00%	3/12/2019 17:23	40.00	ug/L
S190319-ICV1	Nickel KED ICPMS 200.8	ug/L	81.57	102.00%	3/12/2019 17:23	80.00	ug/L
S190319-ICV1	Silver KED ICPMS 200.8	ug/L	40.30	101.00%	3/12/2019 17:23	40.00	ug/L
S190319-ICV1	Zinc KED ICPMS 200.8	ug/L	165.17	103.00%	3/12/2019 17:23	160.00	ug/L
S190319-CCV1	Arsenic KED ICPMS 200.8	ug/L	51.13	102.00%	3/12/2019 17:47	50.00	ug/L
S190319-CCV1	Cadmium KED ICPMS 200.8	ug/L	100.63	101.00%	3/12/2019 17:47	100.00	ug/L
S190319-CCV1	Chromium KED ICPMS 200.8	ug/L	303.13	101.00%	3/12/2019 17:47	300.00	ug/L
S190319-CCV1	Copper KED ICPMS 200.8	ug/L	101.13	101.00%	3/12/2019 17:47	100.00	ug/L
S190319-CCV1	Lead KED ICPMS 200.8	ug/L	51.82	104.00%	3/12/2019 17:47	50.00	ug/L
S190319-CCV1	Nickel KED ICPMS 200.8	ug/L	100.68	101.00%	3/12/2019 17:47	100.00	ug/L
S190319-CCV1	Silver KED ICPMS 200.8	ug/L	50.77	102.00%	3/12/2019 17:47	50.00	ug/L
S190319-CCV1	Zinc KED ICPMS 200.8	ug/L	203.59	102.00%	3/12/2019 17:47	200.00	ug/L
S190319-CCV2	Arsenic KED ICPMS 200.8	ug/L	50.92	102.00%	3/12/2019 18:51	50.00	ug/L
S190319-CCV2	Cadmium KED ICPMS 200.8	ug/L	99.79	99.80%	3/12/2019 18:51	100.00	ug/L
S190319-CCV2	Chromium KED ICPMS 200.8	ug/L	298.53	99.50%	3/12/2019 18:51	300.00	ug/L
S190319-CCV2	Copper KED ICPMS 200.8	ug/L	100.57	101.00%	3/12/2019 18:51	100.00	ug/L
S190319-CCV2	Lead KED ICPMS 200.8	ug/L	49.76	99.50%	3/12/2019 18:51	50.00	ug/L
S190319-CCV2	Nickel KED ICPMS 200.8	ug/L	99.36	99.40%	3/12/2019 18:51	100.00	ug/L
S190319-CCV2	Silver KED ICPMS 200.8	ug/L	49.64	99.30%	3/12/2019 18:51	50.00	ug/L
S190319-CCV2	Zinc KED ICPMS 200.8	ug/L	204.32	102.00%	3/12/2019 18:51	200.00	ug/L
S190319-CCV3	Arsenic KED ICPMS 200.8	ug/L	50.61	101.00%	3/12/2019 20:04	50.00	ug/L
S190319-CCV3	Cadmium KED ICPMS 200.8	ug/L	100.73	101.00%	3/12/2019 20:04	100.00	ug/L
S190319-CCV3	Chromium KED ICPMS 200.8	ug/L	294.06	98.00%	3/12/2019 20:04	300.00	ug/L
S190319-CCV3	Copper KED ICPMS 200.8	ug/L	102.21	102.00%	3/12/2019 20:04	100.00	ug/L
S190319-CCV3	Lead KED ICPMS 200.8	ug/L	49.49	99.00%	3/12/2019 20:04	50.00	ug/L
S190319-CCV3	Nickel KED ICPMS 200.8	ug/L	98.74	98.70%	3/12/2019 20:04	100.00	ug/L
S190319-CCV3	Silver KED ICPMS 200.8	ug/L	50.82	102.00%	3/12/2019 20:04	50.00	ug/L
S190319-CCV3	Zinc KED ICPMS 200.8	ug/L	204.41	102.00%	3/12/2019 20:04	200.00	ug/L
S190319-CCV4	Arsenic KED ICPMS 200.8	ug/L	51.13	102.00%	3/12/2019 21:12	50.00	ug/L
S190319-CCV4	Cadmium KED ICPMS 200.8	ug/L	101.37	101.00%	3/12/2019 21:12	100.00	ug/L
S190319-CCV4	Chromium KED ICPMS 200.8	ug/L	294.58	98.20%	3/12/2019 21:12	300.00	ug/L
S190319-CCV4	Copper KED ICPMS 200.8	ug/L	103.32	103.00%	3/12/2019 21:12	100.00	ug/L
S190319-CCV4	Lead KED ICPMS 200.8	ug/L	49.09	98.20%	3/12/2019 21:12	50.00	ug/L
S190319-CCV4	Nickel KED ICPMS 200.8	ug/L	99.91	99.90%	3/12/2019 21:12	100.00	ug/L
S190319-CCV4	Silver KED ICPMS 200.8	ug/L	51.01	102.00%	3/12/2019 21:12	50.00	ug/L
S190319-CCV4	Zinc KED ICPMS 200.8	ug/L	205.70	103.00%	3/12/2019 21:12	200.00	ug/L
S190319-CCV5	Arsenic KED ICPMS 200.8	ug/L	50.62	101.00%	3/12/2019 22:26	50.00	ug/L
S190319-CCV5	Cadmium KED ICPMS 200.8	ug/L	101.50	102.00%	3/12/2019 22:26	100.00	ug/L
S190319-CCV5	Chromium KED ICPMS 200.8	ug/L	293.69	97.90%	3/12/2019 22:26	300.00	ug/L
S190319-CCV5	Copper KED ICPMS 200.8	ug/L	103.02	103.00%	3/12/2019 22:26	100.00	ug/L
S190319-CCV5	Lead KED ICPMS 200.8	ug/L	49.70	99.40%	3/12/2019 22:26	50.00	ug/L
S190319-CCV5	Nickel KED ICPMS 200.8	ug/L	100.52	101.00%	3/12/2019 22:26	100.00	ug/L
S190319-CCV5	Silver KED ICPMS 200.8	ug/L	51.15	102.00%	3/12/2019 22:26	50.00	ug/L

S190319-CCV5	Zinc KED ICPMS 200.8	ug/L	204.78	102.00%	3/12/2019 22:26	200.00	ug/L
S190319-CCV6	Arsenic KED ICPMS 200.8	ug/L	50.25	101.00%	3/12/2019 23:35	50.00	ug/L
S190319-CCV6	Cadmium KED ICPMS 200.8	ug/L	102.18	102.00%	3/12/2019 23:35	100.00	ug/L
S190319-CCV6	Chromium KED ICPMS 200.8	ug/L	290.74	96.90%	3/12/2019 23:35	300.00	ug/L
S190319-CCV6	Copper KED ICPMS 200.8	ug/L	102.67	103.00%	3/12/2019 23:35	100.00	ug/L
S190319-CCV6	Lead KED ICPMS 200.8	ug/L	49.63	99.30%	3/12/2019 23:35	50.00	ug/L
S190319-CCV6	Nickel KED ICPMS 200.8	ug/L	99.86	99.90%	3/12/2019 23:35	100.00	ug/L
S190319-CCV6	Silver KED ICPMS 200.8	ug/L	51.40	103.00%	3/12/2019 23:35	50.00	ug/L
S190319-CCV6	Zinc KED ICPMS 200.8	ug/L	203.21	102.00%	3/12/2019 23:35	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190321-ICV1	Arsenic KED ICPMS 200.8	39.64	ug/L	99.10%	3/13/2019 15:58	40.00	ug/L
S190321-ICV1	Cadmium KED ICPMS 200.8	78.51	ug/L	98.10%	3/13/2019 15:58	80.00	ug/L
S190321-ICV1	Chromium KED ICPMS 200.8	239.21	ug/L	99.70%	3/13/2019 15:58	240.00	ug/L
S190321-ICV1	Copper KED ICPMS 200.8	81.26	ug/L	102.00%	3/13/2019 15:58	80.00	ug/L
S190321-ICV1	Lead KED ICPMS 200.8	39.86	ug/L	99.70%	3/13/2019 15:58	40.00	ug/L
S190321-ICV1	Nickel KED ICPMS 200.8	80.29	ug/L	100.00%	3/13/2019 15:58	80.00	ug/L
S190321-ICV1	Silver KED ICPMS 200.8	39.50	ug/L	98.80%	3/13/2019 15:58	40.00	ug/L
S190321-ICV1	Zinc KED ICPMS 200.8	164.32	ug/L	103.00%	3/13/2019 15:58	160.00	ug/L
S190321-CCV1	Arsenic KED ICPMS 200.8	50.82	ug/L	102.00%	3/13/2019 16:23	50.00	ug/L
S190321-CCV1	Cadmium KED ICPMS 200.8	99.88	ug/L	99.90%	3/13/2019 16:23	100.00	ug/L
S190321-CCV1	Chromium KED ICPMS 200.8	299.07	ug/L	99.70%	3/13/2019 16:23	300.00	ug/L
S190321-CCV1	Copper KED ICPMS 200.8	100.94	ug/L	101.00%	3/13/2019 16:23	100.00	ug/L
S190321-CCV1	Lead KED ICPMS 200.8	51.01	ug/L	102.00%	3/13/2019 16:23	50.00	ug/L
S190321-CCV1	Nickel KED ICPMS 200.8	100.15	ug/L	100.00%	3/13/2019 16:23	100.00	ug/L
S190321-CCV1	Silver KED ICPMS 200.8	50.27	ug/L	101.00%	3/13/2019 16:23	50.00	ug/L
S190321-CCV1	Zinc KED ICPMS 200.8	204.74	ug/L	102.00%	3/13/2019 16:23	200.00	ug/L
S190321-CCV2	Arsenic KED ICPMS 200.8	50.24	ug/L	100.00%	3/13/2019 17:37	50.00	ug/L
S190321-CCV2	Cadmium KED ICPMS 200.8	99.44	ug/L	99.40%	3/13/2019 17:37	100.00	ug/L
S190321-CCV2	Chromium KED ICPMS 200.8	297.20	ug/L	99.10%	3/13/2019 17:37	300.00	ug/L
S190321-CCV2	Copper KED ICPMS 200.8	101.48	ug/L	101.00%	3/13/2019 17:37	100.00	ug/L
S190321-CCV2	Lead KED ICPMS 200.8	50.51	ug/L	101.00%	3/13/2019 17:37	50.00	ug/L
S190321-CCV2	Nickel KED ICPMS 200.8	99.24	ug/L	99.20%	3/13/2019 17:37	100.00	ug/L
S190321-CCV2	Silver KED ICPMS 200.8	49.60	ug/L	99.20%	3/13/2019 17:37	50.00	ug/L
S190321-CCV2	Zinc KED ICPMS 200.8	204.88	ug/L	102.00%	3/13/2019 17:37	200.00	ug/L
S190321-CCV3	Arsenic KED ICPMS 200.8	50.23	ug/L	100.00%	3/13/2019 18:45	50.00	ug/L
S190321-CCV3	Cadmium KED ICPMS 200.8	99.26	ug/L	99.30%	3/13/2019 18:45	100.00	ug/L
S190321-CCV3	Chromium KED ICPMS 200.8	297.20	ug/L	99.10%	3/13/2019 18:45	300.00	ug/L
S190321-CCV3	Copper KED ICPMS 200.8	101.37	ug/L	101.00%	3/13/2019 18:45	100.00	ug/L
S190321-CCV3	Lead KED ICPMS 200.8	50.75	ug/L	102.00%	3/13/2019 18:45	50.00	ug/L
S190321-CCV3	Nickel KED ICPMS 200.8	99.22	ug/L	99.20%	3/13/2019 18:45	100.00	ug/L
S190321-CCV3	Silver KED ICPMS 200.8	49.63	ug/L	99.30%	3/13/2019 18:45	50.00	ug/L
S190321-CCV3	Zinc KED ICPMS 200.8	205.12	ug/L	103.00%	3/13/2019 18:45	200.00	ug/L
S190321-CCV4	Arsenic KED ICPMS 200.8	50.39	ug/L	101.00%	3/13/2019 19:53	50.00	ug/L
S190321-CCV4	Cadmium KED ICPMS 200.8	99.63	ug/L	99.60%	3/13/2019 19:53	100.00	ug/L
S190321-CCV4	Chromium KED ICPMS 200.8	300.65	ug/L	100.00%	3/13/2019 19:53	300.00	ug/L
S190321-CCV4	Copper KED ICPMS 200.8	102.46	ug/L	102.00%	3/13/2019 19:53	100.00	ug/L
S190321-CCV4	Lead KED ICPMS 200.8	51.16	ug/L	102.00%	3/13/2019 19:53	50.00	ug/L
S190321-CCV4	Nickel KED ICPMS 200.8	101.20	ug/L	101.00%	3/13/2019 19:53	100.00	ug/L
S190321-CCV4	Silver KED ICPMS 200.8	49.76	ug/L	99.50%	3/13/2019 19:53	50.00	ug/L
S190321-CCV4	Zinc KED ICPMS 200.8	206.44	ug/L	103.00%	3/13/2019 19:53	200.00	ug/L
S190321-CCV5	Arsenic KED ICPMS 200.8	50.56	ug/L	101.00%	3/13/2019 21:06	50.00	ug/L
S190321-CCV5	Cadmium KED ICPMS 200.8	98.67	ug/L	98.70%	3/13/2019 21:06	100.00	ug/L
S190321-CCV5	Chromium KED ICPMS 200.8	296.32	ug/L	98.80%	3/13/2019 21:06	300.00	ug/L
S190321-CCV5	Copper KED ICPMS 200.8	102.00	ug/L	102.00%	3/13/2019 21:06	100.00	ug/L
S190321-CCV5	Lead KED ICPMS 200.8	52.01	ug/L	104.00%	3/13/2019 21:06	50.00	ug/L
S190321-CCV5	Nickel KED ICPMS 200.8	99.95	ug/L	100.00%	3/13/2019 21:06	100.00	ug/L
S190321-CCV5	Silver KED ICPMS 200.8	48.99	ug/L	98.00%	3/13/2019 21:06	50.00	ug/L

S190321-CCV5	Zinc KED ICPMS 200.8	206.03	ug/L	103.00%	3/13/2019 21:06	200.00	ug/L
S190321-CCV6	Arsenic KED ICPMS 200.8	50.49	ug/L	101.00%	3/13/2019 22:13	50.00	ug/L
S190321-CCV6	Cadmium KED ICPMS 200.8	99.65	ug/L	99.60%	3/13/2019 22:13	100.00	ug/L
S190321-CCV6	Chromium KED ICPMS 200.8	295.11	ug/L	98.40%	3/13/2019 22:13	300.00	ug/L
S190321-CCV6	Copper KED ICPMS 200.8	101.86	ug/L	102.00%	3/13/2019 22:13	100.00	ug/L
S190321-CCV6	Lead KED ICPMS 200.8	52.19	ug/L	104.00%	3/13/2019 22:13	50.00	ug/L
S190321-CCV6	Nickel KED ICPMS 200.8	100.28	ug/L	100.00%	3/13/2019 22:13	100.00	ug/L
S190321-CCV6	Silver KED ICPMS 200.8	49.76	ug/L	99.50%	3/13/2019 22:13	50.00	ug/L
S190321-CCV6	Zinc KED ICPMS 200.8	205.51	ug/L	103.00%	3/13/2019 22:13	200.00	ug/L
S190321-CCV7	Arsenic KED ICPMS 200.8	50.28	ug/L	101.00%	3/13/2019 23:26	50.00	ug/L
S190321-CCV7	Cadmium KED ICPMS 200.8	99.37	ug/L	99.40%	3/13/2019 23:26	100.00	ug/L
S190321-CCV7	Chromium KED ICPMS 200.8	297.45	ug/L	99.10%	3/13/2019 23:26	300.00	ug/L
S190321-CCV7	Copper KED ICPMS 200.8	101.65	ug/L	102.00%	3/13/2019 23:26	100.00	ug/L
S190321-CCV7	Lead KED ICPMS 200.8	52.22	ug/L	104.00%	3/13/2019 23:26	50.00	ug/L
S190321-CCV7	Nickel KED ICPMS 200.8	100.77	ug/L	101.00%	3/13/2019 23:26	100.00	ug/L
S190321-CCV7	Silver KED ICPMS 200.8	49.54	ug/L	99.10%	3/13/2019 23:26	50.00	ug/L
S190321-CCV7	Zinc KED ICPMS 200.8	205.29	ug/L	103.00%	3/13/2019 23:26	200.00	ug/L

LabNumber	Analysis	Result	Units	Recovery	Analyzed	Spike Conc	Spike Unit
S190324-ICV1	Zinc KED ICPMS 200.8	165.82	ug/L	104.00%	3/14/2019 15:26	160.00	ug/L
S190324-CCV1	Zinc KED ICPMS 200.8	200.74	ug/L	100.00%	3/14/2019 15:50	200.00	ug/L
S190324-CCV2	Zinc KED ICPMS 200.8	198.28	ug/L	99.10%	3/14/2019 16:12	200.00	ug/L